

**Proficiency test SPIL-2
(2015)**

**Nitrogen parameters in wastewater
(untreated sewage)**

Proficiency test SPIL-2 (2015) Quality Documentation

Eurofins Miljø A/S
Smedeskovvej 38
DK-8464 Galten
Denmark

July 2015

Tlf: +45 7022 4266
Fax: +45 7022 4255
e-mail: MBF@eurofins.dk
Web: www.eurofins.dk



Client Environmental laboratories	Client's representative
--	-------------------------

Project Proficiency test SPIL-2 (2015)	Project No 20404-52
---	----------------------------

Authors Maj-Britt Fruekilde	Date 2015-07-16
	Approved by Stine Kjær Ottsen

--	--	--	--	--

--	--	--	--	--

	Quality Documentation Report	<i>MBF</i>	<i>SJN</i>	2015-07-16
--	------------------------------	------------	------------	------------

Revision	Description	By	Approved	Date
----------	-------------	----	----------	------

Key words Analytical quality, assigned value, precision, trueness, homogeneity, stability, total nitrogen, ammonium, nitrite+nitrate, nitrate, conductivity, pH, wastewater	Classification <input checked="" type="checkbox"/> Open <input type="checkbox"/> Internal <input type="checkbox"/> Proprietary
--	---

Distribution DANAK Eurofins:	Maj-Britt Fruekilde, Stine Kjær Ottsen
------------------------------------	--

CONTENTS

1	INTRODUCTION.....	1
2	FEATURES OF THE PROFICIENCY TEST	2
2.1	Sample preparation.....	2
2.2	Statistical analysis of participants' data	2
2.3	Assigned and spike value.....	2
2.3.1	Assigned and spike values.....	3
2.3.2	Test of spike values	3
2.3.3	Test of assigned values.....	3
3	HOMOGENEITY AND STABILITY OF SAMPLES.....	4
4	CONCLUSION	5
5	REFERENCES	6
ANNEX A	LIST OF PARTICIPANTS.....	9
ANNEX B	SAMPLE PREPARATION	11
ANNEX C	CONTROL OF SPIKE VALUES	12
ANNEX D	CONTROL OF RECOVERY.....	18
ANNEX E	CONCENTRATION LEVEL.....	24
ANNEX F	HOMOGENEITY AND STABILITY	25

1 INTRODUCTION

A proficiency test on the analysis of nitrogen parameters in wastewater was conducted on 7 May 2015. The proficiency test was organised by Eurofins Miljø A/S.

The present report contains Eurofins' documentation for the quality of the proficiency test. Results of the proficiency test including data from participating laboratories and statistical analysis of these data were issued in a report to all participants /1/ on 8 June 2015.

2 FEATURES OF THE PROFICIENCY TEST

Participants in the proficiency test were a total of 64 laboratories from Denmark, Norway and Sweden. A list of participants is shown in Appendix A.

The closing date for submission of results was 22 May 2015. All participants except laboratory no. 7, 16, 17 and 22 had submitted their results before the dead-line.

2.1 Sample preparation

The parameters covered in the proficiency test are listed in Table 2 as are the abbreviations used in this report.

Four samples were dispatched for the proficiency test. The samples were sample pairs covering the parameters as described in Table 1. The matrix of the samples represented wastewater, in this case untreated sewage. Sample preparation is described in Appendix B.

Table 1 Samples in the proficiency test

Sample name	Parameters
A1/B1	TN, NH ₄ , NO ₂ +NO ₃ , NO ₃ , γ ₂₅
A2/B2	pH

2.2 Statistical analysis of participants' data

A split-level design was used. The data analysis was performed in accordance with ISO 5725: "Accuracy (trueness and precision) of measurement methods and results" (1994) /2/ and as described in detail in Spliid (1992) /3/. A short introduction to the statistics and a list of symbols and abbreviations used is given in Eurofins document "Schedule for a proficiency test", which is available at Eurofins' home page /4/.

The statistical model used is based on the assumption that the variances for the two samples in a sample pair are identical. The assumption was tested (F-test, 95% confidence level) and the result was that the two variances may be assumed to be identical for all parameters.

2.3 Assigned and spike value

An overview of the concentrations in the samples (the assigned values) and the difference in concentration between the two samples of a sample pair (spike value) are shown in Table 2 compared to the range of concentrations normally encountered in untreated sewage. The table also gives the expanded uncertainty of the assigned values.

Table 2 Assigned and spike value

Parameter	Abbreviation	Unit	Typical Range	Assigned value	Uncertainty of assigned value	Spike value
Total nitrogen	TN	mg/L N	30-80	34	1.1	4
Ammonium	NH ₄	mg/L N	20-50	29.0	0.31	3.2
Nitrite+nitrate	NO ₂₊₃	mg/L N	< 5	0.95	0.068	0.23
Nitrate	NO ₃	mg/L N	< 5	1.09	0.062	0.23
Conductivity	γ_{25}	mS/m	50-2000	88.9	0.59	3.6
pH	pH		6-9	6.92	0.039	0

2.3.1 Assigned and spike values

The content of each parameter in each sample is given an assigned value for the sample with the lower content and a spike value, the spike value being the difference in concentration between the two samples of the sample pair.

In order to ensure optimal use of the data, the assigned value is calculated as the average of the median for both samples in the sample pair after subtraction of the spike value. The spike values are calculated from sample preparation except for γ_{25} where the spike value is the difference between median values for the two samples in the sample pair.

The assigned value for TN is operationally defined and is a consensus value based upon the median for method no. 1, 2, 3, 4, 5. A list of method identification numbers is found in the report to participants /1/. Assigned values for NH₄, NO₂₊₃, NO₃, conductivity and pH are consensus values for all laboratories based on the median. The assigned value for NO₃ (1.09 mg/L N) is higher than the assigned value for NO₂₊₃ (0.95 mg/L N). This is due to differences in results obtained from standard methods and test kit methods.

2.3.2 Test of spike values

A comparison was made (t-test, 95% confidence level) between the spike value and the difference in concentration between the two samples in the sample pair found from the laboratories' results, see Appendix C. The test showed no significant difference between the two.

2.3.3 Test of assigned values

The assigned value and the average of the results obtained from all laboratories were also compared (t-test, 95% confidence level), see Appendix D. The test showed no significant difference between the two and the control of assigned value at Eurofins confirmed the value (Appendix E).

3 **HOMOGENEITY AND STABILITY OF SAMPLES**

The homogeneity and stability of samples were tested using the following parameters as indicators:

NH ₄	Homogeneity test
NO ₃	Combined homogeneity and stability test
pH	Homogeneity test

The results of control measurements are shown in Appendix F. The appendix also gives the results of the statistical evaluation of the control data. The data are analysed by analysis of variance (ANOVA) giving:

1. the standard deviation/variance for replicates (the contribution from analytical variability),
2. the between bottle standard deviation/variance (the contribution from heterogeneity) and
3. the between days concentration difference (the contribution from instability).

Homogeneity is evaluated by comparing the between bottle variance to $0.3 \cdot \sigma$ the standard deviation for evaluation of participants' performance ($0.3 \cdot \sigma$) specified by the Danish EPA /5/, whereas the stability is evaluated by comparing the concentration change of the samples to $0.3 \cdot \sigma$. This test ensures that heterogeneity and instability will not have negative influence on the evaluation of participant performance /6/.

The appendix also shows the standard deviation within and between laboratories from the proficiency test to allow comparison between tests performed and average quality from participating laboratories.

The tests for stability and homogeneity show that the samples are stable and homogeneous.

4 CONCLUSION

The quality control performed, including test of sample stability and homogeneity as well as test of recovery of spike and assigned values, shows that the samples and their assigned values are suitable for testing the proficiency of the participating laboratories for all parameters. The results are also suitable for estimation of the general quality of analyses among all participating laboratories.

5 REFERENCES

- /1/ Eurofins A/S, *Proficiency test SPIL-2 (2015)*, Report to participants, June 2015.
- /2/ ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*, 1994.
- /3/ Spliid, H., *Procedure and analysis of data for proficiency tests and environmental analyses*, Report to Danish Environmental Protection Agency, 1994 (in Danish).
- /4/ Eurofins A/S, *Schedule for a proficiency test*, document may be downloaded from www.eurofins.dk/proficiencytest.
- /5/ Ministry of Environment regulation no. 231 on *quality criteria for environmental measurements*, 5 March 2014 (in Danish).
- /6/ ISO 13528, *Statistical methods for use in proficiency testing by interlaboratory comparison*, 2005.

A N N E X E S

ANNEX A LIST OF PARTICIPANTS

Laboratory	Town	Country
Analyseenheden, Institut for Agroøkologi	Tjele	Denmark
AquaDjurs - Fornæs Renseanlæg	Grenaa	Denmark
Bjergmarken R/A, Roskilde Forsyning	Roskilde	Denmark
Biofos A/S	København K	Denmark
CP Kelco ApS, Spildevandslaboratoriet	Ll. Skensved	Denmark
Esbjerg Forsyning Spildevandslaboratorium	Esbjerg	Denmark
Esbjerg Forsyning Spildevandslaboratorium	Esbjerg	Denmark
Eurofins Miljø A/S	Vejen	Denmark
Faxe Forsyning	Faxe	Denmark
Faxe Forsyning	Faxe	Denmark
FORCE Technology	Holstebro	Denmark
Greve Solrød Forsyning	Greve	Denmark
Halsnæs Kommunale Forsyning A/S	Liseleje	Denmark
Hedensted Spildevand A/S	Daugård	Denmark
Hedensted Spildevand A/S	Daugård	Denmark
Hillerød Forsyning Spildevand A/S	Hillerød	Denmark
Holbæk Forsyning	Holbæk	Denmark
Holstebro Centralrenseanlæg, Vestforsyning A/S	Holstebro	Denmark
Kerteminde Forsyning - Spildevand A/S	Kerteminde	Denmark
Kolding Spildevand A/S	Bjert	Denmark
Middelfart Spildevand A/S	Middelfart	Denmark
Mølleåværkets Driftslaboratorium	Lyngby	Denmark
Nyborg Renseanlæg	Nyborg	Denmark
NK-Spildevand Lab	Næstved	Denmark
Provas Haderslev Forsyningservice A/S	Haderslev	Denmark
Randers Spildevand A/S	Randers SØ	Denmark
Rønne Renseanlæg	Rønne	Denmark
SK Forsyning, Slagelse Renseanlæg	Slagelse	Denmark
SK Forsyning, Slagelse Renseanlæg	Slagelse	Denmark
SK Forsyning, Slagelse Renseanlæg	Slagelse	Denmark
SK Forsyning, Slagelse Renseanlæg	Slagelse	Denmark
SK Forsyning, Slagelse Renseanlæg	Slagelse	Denmark
Svendborg Centralrenseanlæg	Skårup Fyn	Denmark
Vejen Renseanlæg	Vejen	Denmark

Vejle Spildevand A/S	Vejle	Denmark
AIControl Hamar	Hamar	Norway
Eurofins Environment Testing Norway AS	Moss	Norway
NGI Miljølaboratorium	Oslo	Norway
Ernemar Laboratorium	Oskarshamn	Sweden
Eurofins Environment Testing Sweden AB	Lidköping	Sweden
GRYAAB AB	Göteborg	Sweden
Hallsta Pappersbruk	Hallstavik	Sweden
Holmen Paper AB, Bravikens Pappersbruk	Norrköping	Sweden
Kalmar Vatten AB, VA-lab	Kalmar	Sweden
Klippans Reningsverk	Klippan	Sweden
Käppalaverket	Lidingö	Sweden
Laboratoriet vid Smedjeholms avolppsreningsverk	Falkenberg	Sweden
Mjölby Kommun	Mjölby	Sweden
Motala Kommun	Motala	Sweden
Nordic Sugar, Örtofta Sockerbruk	Eslöv	Sweden
NSVA/Öresundsverket	Helsingborg	Sweden
Nynäshamn Kommun, VA-avdelningen	Nynäshamn	Sweden
Preem AB Göteborg	Göteborg	Sweden
Reningsverket Aggerud	Karlskoga	Sweden
Smurfit Kappa Kraftliner	Piteå	Sweden
St1 Refinery AB	Göteborg	Sweden
Södra Cell Mörrum	Mörrum	Sweden
Tekniska förvaltningen, Verksamhetsstöd VA, Laboratoriet	Örebro	Sweden
Uddebo Laboratorium	Luleå	Sweden
Uppsala Vatten o. Avfall AB, Vattenlaboratoriet	Uppsala	Sweden
VIVAB	Falkenberg	Sweden
Västerviks Miljö & Energi AB, Vattenlaboratoriet	Västervik	Sweden
Vallviks Bruk AB	Vallvik	Sweden
Yara AB	Köping	Sweden

ANNEX B SAMPLE PREPARATION

Stock solution	Prepared from	Concentration
Stock TN	25.018 g Disodium edetate, 2H ₂ O milli-Q water up to 1000.0 g	TN: 1882 mg/kg N
Stock NH ₄	10.012 g Ammonium chloride (NH ₄ Cl) milli-Q water up to 1000.0 g	NH ₄ : 2619 mg/kg N
Stock NO ₃	2.502 g Potassium nitrate (KNO ₃) milli-Q water up to 1000.0 g	NO ₃ : 346.4 mg/kg N

Sample	Sample prepared from	TN mg/L N	NH ₄ mg/L N	NO ₃ mg/L N	Conductivity mS/m
A1	75.01 g stock NH ₄ 40.00 g stock NO ₃ 25.00 g stock TN Sample B1 up to 59.99 kg	0.998 · (c+3.14) + 0.785	0.998 · a + 3.279	0.998 · (b+0.867) + 0.231	e
B1	0 g stock NH ₄ 300.02 g stock NO ₃ 200.00 g stock TN Filtered untreated wastewater from Varde sewage treatment plant up to 120.00 kg	c+3.14	a	b+0.867	d

Sample	Sample prepared from	pH
A2/B2	Filtered untreated wastewater from Varde sewage treatment plant.	f

ANNEX C CONTROL OF SPIKE VALUES

Total nitrogen, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB	
1	-1.80	
2	0.70	
3	-1.90	
4A	-0.70	
4B	-2.50	
5	0.70	
6	1.40	
8	-0.53	
9	0.19	
10	0.30	
11	0.10	
12	-0.20	
13	1.20	
14	-1.40	
15	-0.90	
18	-0.50	
19	-0.81	
20A	-0.50	
20B	1.20	
21	-	
23	-0.20	
24	0.10	
25	0.40	
26	0.80	
27	1.10	
28	-1.70	
29	-0.50	
30	-2.00	
31	0.10	
32	-0.10	
33	0.00	
34	-0.25	
35	0.90	
36	-0.24	
37	-0.74	
38	0.90	
39	0.90	
40	-3.00	
41	0.00	
42A	-1.00	
42B	-0.90	
43	0.90	
44	-0.20	
45	0.80	
46	0.16	
47	-2.80	
48	-0.40	
49	-3.20	
50A	1.00	
50B	5.10	UC
51	1.53	

52	0.40	
53	0.20	
54	1.90	
55	-0.10	
56	0.28	
57	-3.50	
58	-2.20	UG
59	0.30	
60	-1.10	
61	0.75	
62	-1.00	
63	-	
64	-0.20	
No of labs., p	60	
No of repl., n	2	
d	-0.26	
s ²	1.44	
s	1.20	
$t = \sqrt{p} \cdot (d/s)$	-1.6845	
Sign. level, p(t)	0.0974	

No test statistics were found to be significant
 UC denotes a Cochran outlier
 UG denotes a Grubbs outlier

Ammonium, mg/L N
Control of differences within sample pairs

Laboratory	Difference AB	
1	1.10	
2	-5.49	UC
3	-0.70	
4A	0.10	
4B	0.20	
5	-1.10	
6	-0.00	
8	-0.10	UG
9	-1.80	
10	-0.70	
11	0.10	
12	-	
13	0.50	
14	0.50	
15	-0.90	
18	0.30	
19	-	
20A	-1.50	
20B	-0.50	
21	-	
23	-0.70	
24	1.00	
25	0.50	
26	-0.23	
27	0.20	
28	-0.60	
29	0.20	
30	0.20	
31	-1.60	
32	-	
33	-0.10	
34	1.20	
35	0.70	
36	0.70	
37	0.00	
38	-	
39	-0.70	
40	-0.20	
41	0.20	
42A	0.10	
42B	-	
43	0.10	
44	0.00	
45	-0.40	
46	-1.33	
47	0.70	
48	-	
49	-2.75	
50A	-0.20	
50B	1.40	
51	-	
52	-2.30	
53	-1.40	
54	0.50	
55	-0.20	

56	0.08
57	-0.20
58	0.10
59	0.10
60	-0.10
61	-0.22
62	-0.85
63	0.24
64	-0.90
No of labs., p	54
No of repl., n	2
d	-0.21
s ²	0.71
s	0.84
t = $\sqrt{p} \cdot (d/s)$	-1.8085
Sign. level, p(t)	0.0762

No test statistics were found to be significant
UC denotes a Cochran outlier
UG denotes a Grubbs outlier

Nitrite+nitrate, mg/L N
Control of differences within sample pairs

Laboratory	Difference AB
1	-0.010
2	-
3	-0.090
4A	-0.030
4B	0.001
5	-0.022
6	-
8	0.010
9	-
10	0.000
11	0.180
12	-
13	-
14	-
15	-
18	-
19	-
20A	-
20B	-
21	-
23	-
24	-
25	0.020
26	0.000
27	-
28	-
29	-
30	0.027
31	-
32	-
33	-
34	-
35	-
36	-
37	0.010
38	0.020
39	0.000
40	-0.130
41	-0.094
42A	0.049
42B	-
43	-
44	-
45	-
46	0.019
47	-
48	-
49	0.010
50A	-0.031
50B	-0.090
51	-
52	-
53	-
54	0.010
55	-
56	0.047

57	-
58	-
59	-0.005
60	-0.010
61	0.039
62	0.110 UG
63	0.013
64	-
No of labs., p	27
No of repl., n	2
d	-0.002
s ²	0.003
s	0.057
t = $\sqrt{p} \cdot (d/s)$	-0.1914
Sign. level, p(t)	0.8497

No test statistics were found to be significant
UG denotes a Grubbs outlier

Nitrate, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB	
1	-	
2	-0.430	UC
3	0.035	
4A	-	
4B	-	
5	0.000	
6	-0.010	
8	0.020	
9	-	
10	0.000	
11	-	
12	-	
13	-0.098	
14	-	
15	-0.010	
18	0.000	
19	-	
20A	-0.320	UC
20B	-0.930	UC
21	-	
23	-0.030	
24	-0.100	
25	-	
26	-	
27	0.000	
28	0.020	
29	0.000	
30	-	
31	-0.030	
32	-	
33	-0.040	
34	-0.090	
35	0.010	
36	-0.110	
37	-	
38	0.030	
39	-	
40	-	
41	-0.088	
42A	0.062	
42B	-	
43	0.020	
44	0.010	
45	-0.030	
46	-	
47	0.020	
48	-	
49	-	
50A	-	
50B	-	
51	-	
52	0.634	UC
53	0.030	
54	-	
55	-0.110	
56	-	

57	0.020
58	-
59	-0.001
60	0.000
61	0.041
62	-
63	-
64	0.060
No of labs., p	32
No of repl., n	2
d	-0.012
s ²	0.002
s	0.049
t = $\sqrt{p} \cdot (d/s)$	-1.3329
Sign. level, p(t)	0.1923

No test statistics were found to be significant
UC denotes a Cochran outlier

Conductivity, mS/m
Control of differences within sample pairs

Laboratory	Difference AB	
1	-0.80	
2	-	
3	-	
4A	28.40	U
4B	-	
5	-	
6	-	
8	-1.71	
9	-	
10	-	
11	0.00	
12	-1.30	UG
13	-0.20	
14	0.20	
15	-	
18	-	
19	-	
20A	36.40	U
20B	46.40	U
21	-0.10	
23	-	
24	-	
25	30.40	U
26	-	
27	-0.30	
28	-	
29	-	
30	-	
31	-	
32	-	
33	-0.20	
34	-	
35	-1.10	
36	-0.30	
37	-	
38	-	
39	31.40	U
40	-	
41	1.00	
42A	0.10	
42B	-	
43	-0.10	
44	-0.10	
45	30.40	U
46	-	
47	-	
48	-	
49	-	
50A	-	
50B	-	
51	-0.20	
52	-	
53	1.80	
54	-	
55	-0.50	
56	-0.28	

57	-	
58	-	
59	0.00	
60	0.50	
61	-0.80	UG
62	-	
63	0.40	
64	-3.30	UC
No of labs., p	21	
No of repl., n	2	
d	-0.09	
s ²	0.50	
s	0.71	
t = $\sqrt{p} \cdot (d/s)$	-0.5844	
Sign. level, p(t)	0.5655	

No test statistics were found to be significant
UC denotes a Cochran outlier
UG denotes a Grubbs outlier
U denotes data excluded from the data file

pH,

Control of differences within sample pairs

Laboratory	Difference AB
1	-0.050
2	-0.079
3	-0.040
4A	-0.070
4B	-
5	-0.020
6	0.300
8	-0.100
9	-0.060
10	-0.110
11	-0.040
12	-0.030
13	-0.010
14	-0.030
15	0.290
18	0.050
19	-0.080
20A	0.370
20B	0.220
21	0.020
23	-0.200
24	-0.070
25	-0.390
26	-0.060
27	-0.070
28	-0.210
29	-0.020
30	0.000
31	0.040
32	-
33	0.000
34	0.010
35	-0.180
36	-0.040
37	-0.020
38	-0.010

39	0.000
40	-0.080
41	-0.030
42A	-0.080
42B	-
43	-0.030
44	-0.040
45	0.000
46	-0.170
47	-0.220
48	-0.100
49	-0.200
50A	-
50B	0.040
51	-
52	-0.310
53	0.030
54	0.040
55	0.010
56	-0.010
57	-0.150
58	0.000
59	0.130
60	0.040
61	-0.040
62	-
63	-0.030
64	0.010
No of labs., p	58
No of repl., n	2
d	-0.032
s ²	0.016
s	0.127
t = $\sqrt{p} \cdot (d/s)$	-1.9353
Sign. level, p(t)	0.0579

No test statistics were found to be significant

ANNEX D CONTROL OF RECOVERY

Total nitrogen, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB	
1	32.60	
2	35.85	
3	36.85	
4A	36.45	
4B	35.25	
5	34.95	
6	34.80	
8	35.51	
9	34.61	
10	40.65	
11	32.75	
12	32.70	
13	33.10	
14	29.40	
15	37.45	
18	34.45	
19	32.25	
20A	34.95	
20B	33.50	
21	-	
23	34.10	
24	32.15	
25	33.60	
26	36.10	
27	30.35	
28	30.55	
29	31.75	
30	30.80	
31	32.45	
32	35.80	
33	31.50	
34	35.17	
35	35.35	
36	34.97	
37	35.70	
38	37.05	
39	33.05	
40	33.70	
41	34.60	
42A	32.50	
42B	32.85	
43	36.65	
44	38.30	
45	34.20	
46	35.55	
47	33.80	
48	32.00	
49	36.15	
50A	26.70	
50B	33.05	UC
51	38.88	

52	33.20	
53	35.80	
54	32.35	
55	33.35	
56	36.85	
57	32.75	
58	22.30	UG
59	33.08	
60	33.65	
61	35.23	
62	30.52	
63	-	
64	34.65	
No of labs., p	60	
No of repl., n	2	
m	34.10	
s ²	5.72	
s	2.39	
Assigned value, μ	34	
Recovery, %	100.3	
$t = \sqrt{p} \cdot (m - \mu) / s$	0.3146	
Sign. level, p(t)	0.7541	

No test statistics were found to be significant
 UC denotes a Cochran outlier
 UG denotes a Grubbs outlier

Ammonium, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB	
1	30.15	
2	31.05	UC
3	28.25	
4A	29.05	
4B	26.50	
5	28.45	
6	30.10	
8	35.85	UG
9	27.82	
10	28.79	
11	28.05	
12	-	
13	29.65	
14	28.45	
15	29.85	
18	29.25	
19	-	
20A	27.35	
20B	27.05	
21	-	
23	29.45	
24	29.70	
25	30.15	
26	28.16	
27	28.10	
28	29.30	
29	30.50	
30	30.10	
31	30.00	
32	-	
33	29.45	
34	30.50	
35	28.65	
36	31.38	
37	28.39	
38	-	
39	28.75	
40	29.30	
41	28.50	
42A	28.95	
42B	-	
43	31.05	
44	29.00	
45	29.10	
46	28.59	
47	29.45	
48	-	
49	30.18	
50A	27.10	
50B	29.20	
51	-	
52	29.75	
53	27.60	
54	27.95	
55	28.90	
56	29.56	

57	25.90
58	28.45
59	29.02
60	30.85
61	28.30
62	31.23
63	28.27
64	28.10
No of labs., p	54
No of repl., n	2
m	28.99
s ²	1.34
s	1.16
Assigned value, μ	29.0
Recovery, %	100.0
$t = \sqrt{p} \cdot (m - \mu) / s$	-0.0417
Sign. level, p(t)	0.9669

No test statistics were found to be significant

UC denotes a Cochran outlier

UG denotes a Grubbs outlier

Nitrite+nitrate, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB
1	0.805
2	-
3	0.865
4A	1.165
4B	0.980
5	1.151
6	-
8	1.105
9	-
10	0.840
11	0.870
12	-
13	-
14	-
15	-
18	-
19	-
20A	-
20B	-
21	-
23	-
24	-
25	1.160
26	0.880
27	-
28	-
29	-
30	0.736
31	-
32	-
33	-
34	-
35	-
36	-
37	0.905
38	0.970
39	1.020
40	0.955
41	0.498
42A	0.756
42B	-
43	-
44	-
45	-
46	0.950
47	-
48	-
49	1.205
50A	0.956
50B	1.285
51	-
52	-
53	-
54	1.095
55	-
56	0.774

57	-
58	-
59	0.907
60	1.095
61	0.908
62	1.725 UG
63	0.923
64	-
No of labs., p	27
No of repl., n	2
m	0.954
s ²	0.029
s	0.171
Assigned value, μ	0.95
Recovery, %	100.4
t = $\sqrt{p} \cdot (m-\mu)/s$	0.1235
Sign. level, p(t)	0.9026

No test statistics were found to be significant
 UG denotes a Grubbs outlier

Nitrate, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB	
1	-	
2	1.205	UC
3	0.962	
4A	-	
4B	-	
5	1.140	
6	1.095	
8	1.090	
9	-	
10	0.830	
11	-	
12	-	
13	0.929	
14	-	
15	1.245	
18	1.120	
19	-	
20A	0.050	UC
20B	0.365	UC
21	-	
23	1.085	
24	1.150	
25	-	
26	-	
27	1.110	
28	1.170	
29	1.120	
30	-	
31	1.195	
32	-	
33	0.800	
34	1.265	
35	1.065	
36	1.465	
37	-	
38	0.955	
39	-	
40	-	
41	0.487	
42A	0.749	
42B	-	
43	1.230	
44	1.105	
45	1.085	
46	-	
47	1.130	
48	-	
49	-	
50A	-	
50B	-	
51	-	
52	0.833	UC
53	0.845	
54	-	
55	1.465	
56	-	

57	1.090
58	-
59	0.893
60	1.080
61	0.893
62	-
63	-
64	1.030
No of labs., p	32
No of repl., n	2
m	1.059
s ²	0.038
s	0.196
Assigned value, μ	1.09
Recovery, %	97.1
t = $\sqrt{p} \cdot (m-\mu)/s$	-0.9107
Sign. level, p(t)	0.3695

No test statistics were found to be significant
UC denotes a Cochran outlier

Conductivity, mS/m

Control of recovery, average of results

Laboratory	Sample pair AB	
1	88.50	
2	-	
3	-	
4A	913.20	U
4B	-	
5	-	
6	-	
8	89.63	
9	-	
10	-	
11	89.20	
12	81.55	UG
13	90.80	
14	88.90	
15	-	
18	-	
19	-	
20A	878.20	U
20B	853.20	U
21	90.85	
23	-	
24	-	
25	921.20	U
26	-	
27	87.45	
28	-	
29	-	
30	-	
31	-	
32	-	
33	89.90	
34	-	
35	87.95	
36	89.35	
37	-	
38	-	
39	907.70	U
40	-	
41	88.40	
42A	90.15	
42B	-	
43	86.75	
44	87.65	
45	909.20	U
46	-	
47	-	
48	-	
49	-	
50A	-	
50B	-	
51	89.70	
52	-	
53	88.30	
54	-	
55	89.35	
56	88.97	

57	-	
58	-	
59	88.80	
60	89.25	
61	79.10	UG
62	-	
63	88.50	
64	83.15	UC
No of labs., p	21	
No of repl., n	2	
m	88.97	
s ²	1.09	
s	1.04	
Assigned value, μ	88.9	
Recovery, %	100.1	
t = $\sqrt{p} \cdot (m-\mu)/s$	0.3041	
Sign. level, p(t)	0.7642	

No test statistics were found to be significant

UC denotes a Cochran outlier

UG denotes a Grubbs outlier

U denotes data excluded from the data file

pH,
Control of recovery, average of results

Laboratory	Sample pair AB
1	6.945
2	6.833
3	7.030
4A	6.775
4B	-
5	6.970
6	7.050
8	7.030
9	6.560
10	6.385
11	6.890
12	6.895
13	6.975
14	6.935
15	6.445
18	6.725
19	6.980
20A	6.325
20B	6.110
21	6.990
23	6.800
24	6.875
25	7.755
26	6.930
27	6.865
28	7.685
29	7.220
30	7.010
31	7.000
32	-
33	6.900
34	6.605
35	6.770
36	6.930
37	6.920
38	7.105
39	6.870
40	6.850
41	7.135
42A	7.030
42B	-
43	7.025
44	6.900
45	6.900
46	6.775
47	7.250
48	7.750
49	6.700
50A	-
50B	6.880
51	7.260
52	7.145
53	6.585
54	6.810
55	7.065
56	7.075

57	6.875
58	6.970
59	6.925
60	7.100
61	7.040
62	-
63	6.925
64	6.725
No of labs., p	58
No of repl., n	2
m	6.923
s ²	0.081
s	0.284
Assigned value, μ	6.92
Recovery, %	100.0
t = $\sqrt{p} \cdot (m-\mu)/s$	0.0756
Sign. level, p(t)	0.9400

No test statistics were found to be significant

ANNEX E CONCENTRATION LEVEL

Parameter	Unit	Sample	Bottle no.	I	II	Bottle Average	Sample	Assigned	Spike	
							Average	value	Measured	Assigned
Total nitrogen	mg/L N	A1	24	38,1	37,0	37,55	37,78	38	4,37	4
			57	37,9	38,3	38,10				
			71	37,8	37,6	37,70				
		B1	26	34,1	33,0	33,55	33,42	34		
			49	33,6	33,3	33,45				
			67	33,4	33,1	33,25				
Ammonium	mg/L N	A1	24	30,0	29,3	29,65	29,88	32,2	2,52	3,2
			57	30,3	29,8	30,05				
			71	29,6	30,3	29,95				
		B1	26	27,8	26,9	27,35	27,37	29,0		
			49	27,3	27,6	27,45				
			67	27,2	27,4	27,30				
Nitrite+nitrate	mg/L N	A1	24	1,29	1,31	1,30	1,32	1,18	0,21	0,23
			57	1,33	1,33	1,33				
			71	1,30	1,33	1,32				
		B1	26	1,07	1,12	1,09	1,10	0,95		
			49	1,08	1,18	1,13				
			67	1,08	1,10	1,09				
Nitrate	mg/L N	A1	24	1,27	1,29	1,28	1,30	1,32	0,22	0,23
			57	1,31	1,31	1,31				
			71	1,28	1,31	1,30				
		B1	26	1,04	1,09	1,07	1,08	1,09		
			49	1,05	1,15	1,10				
			67	1,05	1,07	1,06				
pH		A2	8	6,79	6,80	6,80	6,74	6,92	0,05	0
			40	6,64	6,68	6,66				
			88	6,74	6,79	6,77				
		B2	28	6,60	6,62	6,61	6,69	6,92		
			49	6,75	6,79	6,77				
			50	6,67	6,72	6,70				

ANNEX F HOMOGENEITY AND STABILITY

PT:SPIL-2	
Parameter:NH4	
Unit:mg/L C	
Sigma:1,9931	6,5% of x

Responsible for tests: SMN / IRL

6,5% level or $1,3 \cdot S_{T \max}$

Homogeneity test Date:2015-04-22

Sample	x(a)	x(b)	average	sd	sd^2
A1-4	29,8	30,7	30,3	0,636	0,405
A1-12	30,4	30,7	30,6	0,212	0,05
A1-15	31,0	30,5	30,8	0,354	0,125
A1-29	31,1	30,8	31,0	0,212	0,045
A1-36	30,7	30,4	30,6	0,212	0,045
A1-44	30,7	30,1	30,4	0,424	0,180
A1-54	30,6	30,6	30,6	0,000	0,000
A1-58	30,8	30,2	30,5	0,424	0,180
A1-66	31,9	30,6	31,3	0,919	0,845
A1-76	30,7	31,1	30,9	0,283	0,080
A1-80	30,5	31,0	30,8	0,354	0,125
A1-89	30,6	30,4	30,5	0,141	0,020

For homogeneity

General average (x)	30,66
Sample average sd (s_x)	0,273
Within-sample sd (s_w):	0,418
Between-samples sd (ss):	0
S_L in the Proficiency Test:	1,080
S_R in the Proficiency Test:	1,230

Conclusions

	ss =0	0.3*sigma=0,60
	/x-y/ =0	
Analytical quality	Is $s_w < 0,15 \cdot \sigma$ NO	
Homogeneity:	Is ss < 0.3*sigma? YES	

PT:	SPIL-2
Parameter:	NO3
Unit:	mg/L
Sigma:	0,1 1,3*s_T max

Responsible for tests:

SMN / IRL

6,5% level or 1,3*s_T max

Homogeneity test

Date: 2015-04-22

Sample	x(a)	x(b)	average	sd	sd^2
A1-4	1,30	1,14	1,2	0,113	0,013
A1-12	1,26	1,25	1,3	0,007	0,000
A1-15	1,31	1,35	1,3	0,028	0,001
A1-29	1,29	1,29	1,3	0,000	0,000
A1-36	1,42	1,3	1,4	0,085	0,007
A1-44	1,29	1,18	1,2	0,078	0,006
A1-54	1,27	1,28	1,3	0,007	0,000
A1-58	1,20	1,39	1,3	0,134	0,018
A1-66	1,31	1,31	1,3	0,000	0,000
A1-76	1,28	1,37	1,3	0,064	0,004
A1-80	1,30	1,31	1,3	0,007	0,000
A1-89	1,31	1,31	1,3	0,000	0,000

For homogeneity

General average (x)	1,3
Sample average sd (s _x)	0,041
Within-sample sd (s _w):	0,064
Between-samples sd (ss):	0
S _L in the Proficiency Test:	0,194
S _R in the Proficiency Test:	0,197

Stability test

Date: 2015-05-07

Sample	x(a)	x(b)
A1-24	1,27	1,29
A1-57	1,31	1,31
A1-71	1,28	1,31

For stability

General average (y):	1,295
/x-y/ =	0,0025

Conclusions

ss =	0	0.3*sigma=	0,03
/x-y/ =	0,0025		
Analytical quality	Is s _w < 0,15*sigma	NO	
Homogeneity:	Is ss < 0.3*sigma?	YES	
Stability:	/x-y/ < 0.3*sigma?	YES	

PT:	SPIL-2	
Parameter:	pH	
Unit:	x	
Sigma:	0,065	1,3*Stma

Responsible for tests:

SMN

6,5% level or $1,3 \cdot S_{T \max}$

Homogeneity test

Date: 2015-04-22

Sample	x(a)	x(b)	average	sd	sd^2
A2-3	6,98	7,02	7,0	0,028	0,001
A2-13	6,99	7,05	7,0	0,042	0,00
A2-19	6,99	7,03	7,0	0,028	0,001
A2-23	6,99	7,04	7,0	0,035	0,001
A2-34	7,02	7,07	7,0	0,035	0,001
A2-42	6,99	7,01	7,0	0,014	0,000
A2-45	7,04	7,08	7,1	0,028	0,001
A2-53	7,04	7,07	7,1	0,021	0,000
A2-61	7,02	7,06	7,0	0,028	0,001
A2-70	7,02	7,04	7,0	0,014	0,000
A2-74	7,03	7,06	7,0	0,021	0,000
A2-86	7,02	7,04	7,0	0,014	0,000

For homogeneity

General average (x)	7,03
Sample average sd (s _x)	0,020
Within-sample sd (s _w):	0,027
Between-samples sd (ss):	0,0065
S _L in the Proficiency Test:	0,277
S _R in the Proficiency Test:	0,291

Conclusions

$ss = 0,007$		$0.3 \cdot \sigma$ $= 0,02$
$ x-y = 0$		
Analytical quality	Is $s_w < 0,15 \cdot \sigma$	NO
Homogeneity:	Is $ss < 0.3 \cdot \sigma$?	YES