

**Proficiency test SPIL-4  
(2015)**

**Nitrogen parameters in wastewater (effluent)**



# Proficiency test SPIL-4 (2015) Quality Documentation

Eurofins Miljø A/S  
Smedeskovvej 38  
DK-8464 Galten  
Denmark

February 2016

Tlf: +45 7022 4266  
Fax: +45 7022 4255  
e-mail: mbf@eurofins.dk  
Web: www.eurofins.dk

Client  Environmental laboratories	Client's representative
--	-------------------------

Project  Proficiency test SPIL-4 (2015)	Project No  20404-54
---	----------------------------

Authors  Maj-Britt Fruekilde	Date 2016-01-29
	Approved by  Ulla Lund

--	--	--	--	--

--	--	--	--	--

	Quality Documentation Report	MBF		2016-02-08
--	------------------------------	-----	---	------------

Revision	Description	By	Approved	Date
----------	-------------	----	----------	------

Key words Analytical quality, assigned value, precision, trueness, homogeneity, stability, total nitrogen, ammonium, nitrite+nitrate, nitrate, conductivity,pH, wastewater	Classification <input checked="" type="checkbox"/> Open <input type="checkbox"/> Internal <input type="checkbox"/> Proprietary
---	---

Distribution DANAK Eurofins:	Maj-Britt Fruekilde, Ulla Lund
------------------------------------	--------------------------------



## **CONTENTS**

1	INTRODUCTION.....	1
2	FEATURES OF THE PROFICIENCY TEST .....	2
2.1	Sample preparation .....	2
2.2	Statistical analysis of participants' data .....	2
2.3	Assigned and spike value.....	2
2.3.1	Assigned and spike values.....	3
2.3.2	Test of spike values .....	4
2.3.3	Test of assigned values.....	4
3	HOMOGENEITY AND STABILITY OF SAMPLES.....	6
4	CONCLUSION .....	7
5	REFERENCES .....	8
ANNEX A	LIST OF PARTICIPANTS.....	10
ANNEX B	SAMPLE PREPARATION .....	11
ANNEX C	CONTROL OF SPIKE VALUES .....	14
ANNEX D	CONTROL OF RECOVERY.....	15
ANNEX E	CONCENTRATION LEVEL.....	16
ANNEX F	HOMOGENEITY AND STABILITY .....	17



## **1 INTRODUCTION**

A proficiency test on the analysis of nitrogen parameters in wastewater was conducted on 12 November 2015. The proficiency test was organised by Eurofins Miljø A/S.

The present report contains Eurofins' documentation for the quality of the proficiency test. Results of the proficiency test including data from participating laboratories and statistical analysis of these data were issued in a report to all participants /1/ on 14 December 2015.

## 2 FEATURES OF THE PROFICIENCY TEST

Participants in the proficiency test were a total of 63 laboratories from Denmark, Norway and Sweden. A list of participants is shown in Appendix A.

The closing date for submission of results was 27 November 2015. All participants except laboratory no. 40 had submitted their results before the dead-line.

### 2.1 Sample preparation

The parameters covered in the proficiency test are listed in Table 2 as are the abbreviations used in this report.

Four samples were dispatched for the proficiency test. The samples were sample pairs covering the parameters as described in Table 1. The matrix of the samples represented wastewater, in this case effluent. Sample preparation is described in Appendix B.

Table 1 Samples in the proficiency test

Sample name	Parameters
A1/B1	TN, NH <sub>4</sub> , NO <sub>2</sub> +NO <sub>3</sub> , NO <sub>3</sub> , γ <sub>25</sub>
A2/B2	pH

### 2.2 Statistical analysis of participants' data

A split-level design was used. The data analysis was performed in accordance with ISO 5725: "Accuracy (trueness and precision) of measurement methods and results" (1994) /2/ and as described in detail in Spliid (1992) /3/. A short introduction to the statistics and a list of symbols and abbreviations used is given in Eurofins document "Schedule for a proficiency test", which is available at Eurofins' home page /4/.

The statistical model used is based on the assumption that the variances for the two samples in a sample pair are identical. The assumption was tested (F-test, 95% confidence level) and the result was that the two variances may be assumed to be identical for all parameters.

### 2.3 Assigned and spike value

An overview of the concentrations in the samples (the assigned values) and the difference in concentration between the two samples of a sample pair (spike value) are shown in Table 2 compared to the range of concentrations normally encountered in effluent. The table also gives the expanded uncertainty of the assigned values.



Table 2 Assigned and spike value

Parameter	Abbreviation	Unit	Typical Range	Assigned value	Uncertainty of assigned value	Spike value
Total nitrogen	TN	mg/L N	2-10	4.0	0.10	0.8
Ammonium	NH <sub>4</sub>	mg/L N	0.1-2	0.567	0.0097	0.130
Nitrite+nitrate	NO <sub>2+3</sub>	mg/L N	1-5	2.07	0.043	0.57
Nitrate	NO <sub>3</sub>	mg/L N	1-5	2.13	0.038	0.57
Conductivity	$\gamma_{25}$	mS/m	50-300	54.2	0.39	0.4
pH	pH		6-9	7.24	0.051	0.00

### 2.3.1 Assigned and spike values

The content of each parameter in each sample is given an assigned value for the sample with the lower content and a spike value, the spike value being the difference in concentration between the two samples of the sample pair.

In order to ensure optimal use of the data, the assigned value is calculated as the average of the median for both samples in the sample pair after subtraction of the spike value. The spike values are calculated from sample preparation except for  $\gamma_{25}$  where the spike value is the difference between adjusted median values for the two samples in the sample pair.

The assigned value for TN is operationally defined and is a consensus value based upon the median for method no. 1, 2, 3, 4, 5 and 7. A list of method identification numbers is found in the report to participants /1/. Assigned values for NH<sub>4</sub>, NO<sub>2+3</sub>, NO<sub>3</sub>, conductivity and pH are consensus values for all laboratories based on the median. The assigned value for NO<sub>3</sub> (2.13 mg/L N) is higher than the assigned value for NO<sub>2+3</sub> (2.07 mg/L N). This is due to differences in results obtained from standard methods and test kit methods.

### 2.3.2 Test of spike values

A comparison was made (t-test, 95% confidence level) between the spike value and the difference in concentration between the two samples in the sample pair found from the laboratories' results, see Appendix C. The test showed no significant difference between the two for most of the parameters. The test revealed a significant difference between the two for pH. However, the difference is numerically small and has insignificant influence on the general quality of analyses estimated from the data as well as on the evaluation of accuracy of participating laboratories.

### 2.3.3 Test of assigned values

The assigned value and the average of the results obtained from all laboratories were also compared (t-test, 95% confidence level), see Appendix D. The test showed no significant difference between the two and the control of assigned value at Eurofins confirmed the value (Appendix E).

### 3 **HOMOGENEITY AND STABILITY OF SAMPLES**

The homogeneity and stability of samples were tested using the following parameters as indicators:

NH <sub>4</sub>	Homogeneity test
NO <sub>3</sub>	Combined homogeneity and stability test
pH	Homogeneity test

The results of control measurements are shown in Appendix F. The appendix also gives the results of the statistical evaluation of the control data. The data are analysed by analysis of variance (ANOVA) giving:

1. the standard deviation/variance for replicates (the contribution from analytical variability),
2. the between bottle standard deviation/variance (the contribution from heterogeneity) and
3. the between days concentration difference (the contribution from instability).

Homogeneity is evaluated by comparing the between bottle variance to  $0.3 \cdot \hat{\sigma}$  the standard deviation for evaluation of participants' performance ( $0.3 \cdot \hat{\sigma}$ ) specified by the Danish EPA /5/, whereas the stability is evaluated by comparing the concentration change of the samples to  $0.3 \cdot \hat{\sigma}$ . This test ensures that heterogeneity and instability will not have negative influence on the evaluation of participant performance /6/.

The appendix also shows the standard deviation within and between laboratories from the proficiency test to allow comparison between tests performed and average quality from participating laboratories.

The tests for stability and homogeneity show that the samples are stable and homogeneous.

## **4 CONCLUSION**

The quality control performed, including test of sample stability and homogeneity as well as test of recovery of spike and assigned values, shows that the samples and their assigned values are suitable for testing the proficiency of the participating laboratories for all parameters. The results are also suitable for estimation of the general quality of analyses among all participating laboratories.

For pH the participants could not recover the spike value. The difference between the calculated spike value and that found by the participants is small and the influence on evaluation of participant performance or estimation of general quality of analyses is insignificant.

## 5 REFERENCES

- /1/ Eurofins A/S, *Proficiency test SPIL-4 (2015)*, Report to participants, December 2015.
- /2/ ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*, 1994.
- /3/ Spliid, H., *Procedure and analysis of data for proficiency tests and environmental analyses*, Report to Danish Environmental Protection Agency, 1994 (in Danish).
- /4/ Eurofins A/S, *Schedule for a proficiency test*, document may be downloaded from [www.eurofins.dk/proficiencytest](http://www.eurofins.dk/proficiencytest).
- /5/ Ministry of Environment regulation no. 231 on *quality criteria for environmental measurements*, 5 March 2014 (in Danish).
- /6/ ISO 13528, *Statistical methods for use in proficiency testing by interlaboratory comparison*, 2005.

***A N N E X E S***



## **ANNEX A LIST OF PARTICIPANTS**

<b>Laboratory</b>	<b>Town</b>	<b>Country</b>
AquaDjurs - Fornæs Renseanlæg	Grenaa	Denmark
Biofos A/S	København K	Denmark
Bjergmarken R/A, Roskilde Forsyning	Roskilde	Denmark
CP Kelco ApS, Spildevandslaboratoriet	Ll. Skensved	Denmark
Esbjerg Forsyning Spildevandslaboratorium	Esbjerg	Denmark
Eurofins Miljø A/S	Vejen	Denmark
Faxe Forsyning	Faxe	Denmark
FORCE Technology	Holstebro	Denmark
Frederikssund Forsyning	Frederikssund	Denmark
Greve Solrød Forsyning	Greve	Denmark
Halsnæs Spildevand A/S	Liseleje	Denmark
Hedensted Spildevand A/S	Daugård	Denmark
Hillerød Forsyning Spildevand A/S	Hillerød	Denmark
Holbæk Forsyning	Holbæk	Denmark
Holstebro Centralrenseanlæg, Vestforsyning A/S	Holstebro	Denmark
Højmarklaboratoriet	Lem St.	Denmark
Kerteminde Forsyning - Spildevand A/S	Kerteminde	Denmark
Kolding Spildevand A/S	Bjert	Denmark
Køge-Egnens Renseanlæg	Køge	Denmark
Middelfart Spildevand A/S	Middelfart	Denmark
Mølleåværkets Driftslaboratorium	Lyngby	Denmark
NK-Spildevand Lab	Næstved	Denmark
Nyborg Renseanlæg	Nyborg	Denmark
Provas Haderslev Forsyningservice A/S	Haderslev	Denmark
Randers Spildevand A/S	Randers SØ	Denmark
Rønne Renseanlæg	Rønne	Denmark
SK Forsyning, Slagelse Renseanlæg	Slagelse	Denmark
Svendborg Centralrenseanlæg	Skårup Fyn	Denmark
Tårnby Forsyning	Kastrup	Denmark
Vandrens - Stignæs Industripark A/S	Skælskør	Denmark
Vejen Renseanlæg	Vejen	Denmark
Vejle Spildevand A/S	Vejle	Denmark
NGI Miljølaboratorium	Oslo	Norway

<b>Laboratory</b>	<b>Town</b>	<b>Country</b>
Ernemar Laboratorium	Oskarshamn	Sweden
Eurofins Environment Testing Sweden AB	Lidköping	Sweden
GRYAAB AB	Göteborg	Sweden
Holmen Paper AB Bravikens Pappersbruk	Norrköping	Sweden
Kalmar Vatten AB, VA-lab	Kalmar	Sweden
Kristianstad Kommun	Kristianstad	Sweden
Käppalaverket	Lidingö	Sweden
Mjölby Kommun	Mjölby	Sweden
Motala Kommun	Motala	Sweden
NSVA/Öresundsverket	Helsingborg	Sweden
Preem AB Göteborg	Göteborg	Sweden
Preemraff Lysekil	Lysekil	Sweden
Reningsverket Aggerud	Karlskoga	Sweden
Smurfit Kappa Kraftliner	Piteå	Sweden
St1 Refinery AB	Göteborg	Sweden
Stora Enso Paper	Hyltebruk	Sweden
Södra Cell AB Mönsterås	Mönsterås	Sweden
Tekniska förvaltningen, Verksamhetsstöd VA, Laboratoriet	Örebro	Sweden
Uppsala Vatten o. Avfall AB, Vattenlaboratoriet	Uppsala	Sweden
Vallviks Bruk AB	Vallvik	Sweden
VIVAB	Falkenberg	Sweden
Västerviks Miljö & Energi AB, Vattenlaboratoriet	Västervik	Sweden
Yara AB	Köping	Sweden



## ANNEX B SAMPLE PREPARATION

Stock solution	Prepared from	Concentration
Stock TN	2.9997 g Disodium edetate, 2H <sub>2</sub> O milli-Q water up to 1000.0 g	TN: 225.8 mg/kg N
Stock NH <sub>4</sub>	1. 0003 g Ammonium chloride (NH <sub>4</sub> Cl) milli-Q water up to 1000.0 g	NH <sub>4</sub> : 262.0 mg/kg N
Stock NO <sub>3</sub>	2.502 g Potassium nitrate (KNO <sub>3</sub> ) milli-Q water up to 1000.0 g	NO <sub>3</sub> : 346.7 mg/kg N

Sample	Sample prepared from	TN mg/L N	NH <sub>4</sub> mg/L N	NO <sub>3</sub> mg/L N	Conductivity mS/m
A1	100.00 g stock NH <sub>4</sub> 150.00 g stock NO <sub>3</sub> 250.00 g stock TN Filtered treated wastewater from Bramming sewage treatment plant up to 55.00 kg	c+2.448	a+0.476	b+0.946	d
B1	15.04 g stock NH <sub>4</sub> 50.02 g stock NO <sub>3</sub> 10.03 g stock TN Sample A1 up to 30.00 kg	0.998 · (c+2.448) + 0.785	0.998 · (a+0.476) + 0.131	0.998 · (b+0.946) + 0.578	e

Sample	Sample prepared from	pH
A2/B2	Filtered treated wastewater from Bramming sewage treatment plant.	f

## ANNEX C CONTROL OF SPIKE VALUES

Total nitrogen, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	-0.350
2	-0.190
3	0.730
4	-1.100
5	-0.010
6	-
7	-0.030
8	-0.040
9	0.200
10	0.040
11	-0.100
12	-0.310
13A	-0.234
14	0.800
15	0.120
16	0.080
17	-0.410
18	-0.200
19	0.250
20	0.110
21	0.020
22	-0.030
23	-0.190
24	0.390
25	0.300
26A	-0.098
26B	-0.060
27	-0.110
28	0.350
29	0.010
30	0.160
31	-0.320
32	-0.020
33	-0.070
34	0.071
35	0.340
36	-0.070
37	-0.170
38	-0.210
39	0.200
41	-0.140
42	0.260
43	-0.380
44	0.140
45	0.191
46	-0.370
47	0.040
48A	0.032
48B	1.020
49A	-0.040
49B	0.210

Laboratory	Difference AB
50	0.040
51	0.290
52A	-0.050
52B	0.060
53	1.060
54	0.135
55	0.870
56	0.220
57	0.180
58	-
59	-0.530
60	-0.603 UG
61	0.130
62	0.100
63	0.050
13B	-
No of labs., p	63
No of repl., n	2
d	0.053
s <sup>2</sup>	0.122
s	0.349
t = $\sqrt{p} \cdot (d/s)$	1.2146
Sign. level, p(t)	0.2291

No test statistics were found to be significant  
UG denotes a Grubbs outlier

Ammonium, mg/L N  
Control of differences within sample pairs

Laboratory	Difference AB
1	0.0000
2	-0.0180
3	0.0100
4	-0.0100
5	0.0120
6	-
7	-0.0070
8	-
9	0.0100
10	-
11	-0.0060
12	-
13A	0.0050
14	-0.0020
15	-0.0100
16	0.0400
17	-0.0130
18	0.0280
19	-0.0130
20	-
21	0.0000
22	0.0110
23	-0.0100
24	-0.0100
25	0.0300
26A	-0.0080
26B	-
27	0.0100
28	-0.0140
29	0.0000
30	0.0200
31	0.0110
32	0.0070
33	-0.0200
34	0.0100
35	-0.0050
36	-0.0100
37	-0.0180
38	0.1010
39	0.1200
41	0.0000
42	-0.0040
43	-0.0200
44	-0.0120
45	-
46	-0.0040
47	0.0000
48A	-0.0035
48B	-
49A	-0.0090
49B	0.0010
50	0.0070
51	-0.0070
52A	-0.0250
52B	-0.0100
53	0.0100

UC  
UC

Laboratory	Difference AB
54	0.0240
55	0.0550
56	-0.0150
57	0.0140
58	0.0070
59	-0.0060
60	-0.0020
61	-0.0170
62	-0.0010
63	-0.0300
13B	-0.0027
No of labs., p	57
No of repl., n	2
d	-0.0004
s <sup>2</sup>	0.0002
s	0.0156
t = $\sqrt{p} \cdot (d/s)$	-0.1711
Sign. level, p(t)	0.8648

No test statistics were found to be significant  
UC denotes a Cochran outlier

Nitrite+Nitrate, mg/L N  
Control of differences within sample pairs

Laboratory	Difference AB
1	0.010
2	-
3	-
4	-
5	-
6	-
7	0.050
8	-
9	-
10	-
11	-0.000
12	-
13A	-0.032
14	-
15	-0.000
16	-
17	-
18	-
19	-
20	-
21	-
22	-
23	-
24	-
25	-
26A	-0.053
26B	-
27	-
28	-0.120
29	-
30	0.040
31	0.010
32	-
33	0.020
34	0.002
35	-
36	0.060
37	-
38	0.150 UG
39	-0.030
41	-
42	-
43	-
44	-
45	-
46	-
47	0.080
48A	-0.045
48B	-
49A	-
49B	-
50	-0.040
51	-
52A	-0.010
52B	0.030
53	-

Laboratory	Difference AB
54	1.260 UC
55	-
56	-
57	-
58	-0.020
59	-
60	-
61	-0.010
62	-
63	-
13B	-
No of labs., p	20
No of repl., n	2
d	-0.003
s <sup>2</sup>	0.002
s	0.045
t = $\sqrt{p} \cdot (d/s)$	-0.2866
Sign. level, p(t)	0.7776

No test statistics were found to be significant  
UC denotes a Cochran outlier  
UG denotes a Grubbs outlier

Nitrate, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	0.010
2	0.020
3	0.000
4	-0.010
5	-0.020
6	-
7	-
8	-
9	-
10	-
11	-0.010
12	-
13A	-0.032
14	0.020
15	-
16	0.060
17	-0.010
18	-0.020
19	-0.050
20	-
21	0.010
22	-
23	0.010
24	0.140
25	0.070
26A	-
26B	-
27	-0.010
28	-
29	-0.030
30	-
31	0.010
32	-0.070
33	-
34	-
35	-0.040
36	0.060
37	0.110
38	0.020
39	-
41	-
42	0.010
43	0.000
44	0.100 UG
45	-
46	-0.010
47	-
48A	-
48B	-
49A	-0.020
49B	0.050
50	-0.040
51	0.010
52A	-
52B	-
53	0.030

Laboratory	Difference AB
54	-
55	0.020
56	-0.020
57	0.020
58	-
59	0.090
60	-0.047
61	-
62	0.030
63	-0.020
13B	-
No of labs., p	39
No of repl., n	2
d	0.009
s <sup>2</sup>	0.002
s	0.044
t = $\sqrt{p} \cdot (d/s)$	1.2454
Sign. level, p(t)	0.2206

No test statistics were found to be significant  
 UG denotes a Grubbs outlier

Conductivity, mS/m  
Control of differences within sample pairs

Laboratory	Difference AB	
1	-	
2	0.40	
3	-2.60	UC
4	-	
5	-0.10	
6	-0.40	
7	0.00	
8	-	
9	-	
10	-0.60	
11	-0.20	
12	-	
13A	-0.15	
14	-	
15	0.34	UG
16	-	
17	-	
18	0.10	
19	-	
20	-	
21	-0.30	
22	-	
23	-	
24	-	
25	1.40	UG
26A	-0.10	
26B	-	
27	-	
28	-	
29	-0.20	
30	-	
31	0.00	
32	-0.70	
33	-	
34	-	
35	0.40	
36	0.90	
37	-	
38	-	
39	0.00	
41	0.10	
42	0.20	
43	-	
44	-0.40	
45	-	
46	-	
47	-	
48A	-	
48B	-	
49A	-	
49B	-	
50	-0.30	
51	-	
52A	-	
52B	-	
53	-	

Laboratory	Difference AB
54	-0.10
55	0.30
56	-
57	-
58	0.20
59	0.50
60	0.50
61	-
62	-
63	-0.10
13B	-
No of labs., p	26
No of repl., n	2
d	-0.00
s <sup>2</sup>	0.13
s	0.36
t = $\sqrt{p} \cdot (d/s)$	-0.0269
Sign. level, p(t)	0.9787

No test statistics were found to be significant  
UC denotes a Cochran outlier  
UG denotes a Grubbs outlier

pH,  
Control of differences within sample pairs

Laboratory	Difference AB
1	0.140
2	-0.100
3	-0.060
4	0.015
5	-0.110
6	-0.070
7	-0.030
8	-
9	-0.080
10	-0.080
11	0.020
12	-0.170
13A	0.000
14	-0.200
15	-0.050
16	-0.130
17	-0.040
18	0.040
19	-0.020
20	-
21	0.010
22	0.030
23	0.050
24	-0.130
25	-0.100
26A	0.020
26B	-
27	-0.090
28	0.000
29	-0.050
30	-0.070
31	-0.090
32	0.040
33	0.000
34	-0.080
35	0.000
36	0.060
37	-0.074
38	-0.010
39	0.020
41	-0.030
42	-0.030
43	-0.040
44	0.000

UG

Laboratory	Difference AB
45	-
46	-0.330
47	-0.030
48A	-0.030
48B	-0.060
49A	-0.030
49B	-
50	-0.070
51	-0.070
52A	-
52B	0.080
53	0.010
54	-0.050
55	0.000
56	0.010
57	0.080
58	-0.210
59	0.010
60	0.090
61	-0.239
62	-0.020
63	-0.060
13B	-
No of labs., p	57
No of repl., n	2
d	-0.036
s <sup>2</sup>	0.005
s	0.073
t = $\sqrt{p} \cdot (d/s)$	-3.6568
Sign. level, p(t)	0.0006 ***

UC

UG

\* denotes that there is a significant difference (t-test, 5%-level)  
 \*\* denotes that there is a significant difference (t-test, 1%-level)  
 \*\*\* denotes that there is a significant difference (t-test, 0.1%-level)  
 UC denotes a Cochran outlier  
 UG denotes a Grubbs outlier

Difference for sample pair AB is significantly different from 0, and data should be corrected with the difference (in spike value), during execution of Cochran's test.

## ANNEX D CONTROL OF RECOVERY

Total nitrogen, mg/L N

Control of differences within sample pairs

Laboratory	Difference AB
1	-0.350
2	-0.190
3	0.730
4	-1.100
5	-0.010
6	-
7	-0.030
8	-0.040
9	0.200
10	0.040
11	-0.100
12	-0.310
13A	-0.234
14	0.800
15	0.120
16	0.080
17	-0.410
18	-0.200
19	0.250
20	0.110
21	0.020
22	-0.030
23	-0.190
24	0.390
25	0.300
26A	-0.098
26B	-0.060
27	-0.110
28	0.350
29	0.010
30	0.160
31	-0.320
32	-0.020
33	-0.070
34	0.071
35	0.340
36	-0.070
37	-0.170
38	-0.210
39	0.200
41	-0.140
42	0.260
43	-0.380
44	0.140
45	0.191
46	-0.370
47	0.040
48A	0.032
48B	1.020
49A	-0.040
49B	0.210

Laboratory	Difference AB
50	0.040
51	0.290
52A	-0.050
52B	0.060
53	1.060
54	0.135
55	0.870
56	0.220
57	0.180
58	-
59	-0.530
60	-0.603 UG
61	0.130
62	0.100
63	0.050
13B	-
No of labs., p	63
No of repl., n	2
d	0.053
s <sup>2</sup>	0.122
s	0.349
t = $\sqrt{p} \cdot (d/s)$	1.2146
Sign. level, p(t)	0.2291

No test statistics were found to be significant  
UG denotes a Grubbs outlier



Ammonium, mg/L N  
Control of recovery, average of results

Laboratory	Sample pair AB
1	0.5700
2	0.5770
3	0.5450
4	0.5870
5	0.5290
6	-
7	0.5625
8	-
9	0.5750
10	-
11	0.5410
12	-
13A	0.5185
14	0.5410
15	0.5870
16	0.5000
17	0.5905
18	0.6230
19	0.5705
20	-
21	0.5460
22	0.5245
23	0.5250
24	0.5690
25	0.5850
26A	0.5250
26B	-
27	0.5600
28	0.5820
29	0.5510
30	0.6770
31	0.6285
32	0.5255
33	0.5900
34	0.5920
35	0.5645
36	0.5750
37	0.5850
38	0.5635 UC
39	0.6400 UC
41	0.5270
42	0.5700
43	0.5900
44	0.5700
45	-
46	0.5540
47	0.5720
48A	0.6196
48B	-
49A	0.5895
49B	0.5465
50	0.5465
51	0.5745
52A	0.5515
52B	0.5650
53	0.5230

Laboratory	Sample pair AB
54	0.5240
55	0.5155
56	0.5705
57	0.5070
58	0.6045
59	0.5410
60	0.5120
61	0.5785
62	0.5525
63	0.6510
13B	0.5213
No of labs., p	57
No of repl., n	2
m	0.5631
s <sup>2</sup>	0.0013
s	0.0356
Assigned value, μ	0.567
Recovery, %	99.3
$t = \sqrt{p} \cdot (m - \mu) / s$	-0.8179
Sign. level, p(t)	0.4169

No test statistics were found to be significant  
UC denotes a Cochran outlier

Nitrite+Nitrate, mg/L N  
Control of recovery, average of results

Laboratory	Sample pair AB
1	2.125
2	-
3	-
4	-
5	-
6	-
7	1.935
8	-
9	-
10	-
11	2.050
12	-
13A	2.072
14	-
15	2.150
16	-
17	-
18	-
19	-
20	-
21	-
22	-
23	-
24	-
25	-
26A	2.027
26B	-
27	-
28	2.210
29	-
30	2.050
31	2.165
32	-
33	2.240
34	2.027
35	-
36	1.950
37	-
38	1.495 UG
39	2.055
41	-
42	-
43	-
44	-
45	-
46	-
47	2.230
48A	2.129
48B	-
49A	-
49B	-
50	2.010
51	-
52A	2.065
52B	2.175
53	-

Laboratory	Sample pair AB
54	2.583 UC
55	-
56	-
57	-
58	2.070
59	-
60	-
61	2.225
62	-
63	-
13B	-
No of labs., p	20
No of repl., n	2
m	2.098
s <sup>2</sup>	0.008
s	0.091
Assigned value, μ	2.07
Recovery, %	101.4
$t = \sqrt{p} \cdot (m-\mu)/s$	1.3777
Sign. level, p(t)	0.1843

No test statistics were found to be significant  
UC denotes a Cochran outlier  
UG denotes a Grubbs outlier

Nitrate, mg/L N

Control of recovery, average of results

Laboratory	Sample pair AB
1	2.125
2	2.010
3	2.230
4	2.125
5	2.030
6	-
7	-
8	-
9	-
10	-
11	2.035
12	-
13A	2.071
14	2.040
15	-
16	2.120
17	2.225
18	2.280
19	2.195
20	-
21	2.135
22	-
23	1.945
24	2.240
25	2.165
26A	-
26B	-
27	2.205
28	-
29	2.075
30	-
31	2.165
32	2.045
33	-
34	-
35	2.140
36	1.950
37	2.085
38	2.050
39	-
41	-
42	2.165
43	2.230
44	2.580 UG
45	-
46	2.075
47	-
48A	-
48B	-
49A	2.200
49B	2.125
50	2.010
51	2.215
52A	-
52B	-
53	1.975

Laboratory	Sample pair AB
54	-
55	2.190
56	2.210
57	2.160
58	-
59	2.045
60	1.927
61	-
62	2.175
63	1.780
13B	-
No of labs., p	39
No of repl., n	2
m	2.107
s <sup>2</sup>	0.011
s	0.106
Assigned value, μ	2.13
Recovery, %	98.9
$t = \sqrt{p} \cdot (m-\mu)/s$	-1.3673
Sign. level, p(t)	0.1796

No test statistics were found to be significant  
 UG denotes a Grubbs outlier

Conductivity, mS/m  
Control of recovery, average of results

Laboratory	Sample pair AB	
1	-	
2	54.90	
3	56.30	UC
4	-	
5	54.55	
6	54.90	
7	52.40	
8	-	
9	-	
10	55.20	
11	55.00	
12	-	
13A	53.98	
14	-	
15	5.35	UG
16	-	
17	-	
18	51.25	
19	-	
20	-	
21	54.65	
22	-	
23	-	
24	-	
25	549.30	UG
26A	53.95	
26B	-	
27	-	
28	-	
29	54.60	
30	-	
31	53.60	
32	57.55	
33	-	
34	-	
35	49.80	
36	54.75	
37	-	
38	-	
39	54.60	
41	53.45	
42	50.90	
43	-	
44	54.20	
45	-	
46	-	
47	-	
48A	-	
48B	-	
49A	-	
49B	-	
50	54.05	
51	-	
52A	-	
52B	-	
53	-	

Laboratory	Sample pair AB
54	54.45
55	50.15
56	-
57	-
58	53.90
59	53.65
60	55.25
61	-
62	-
63	53.05
13B	-
No of labs., p	26
No of repl., n	2
m	53.80
s <sup>2</sup>	2.92
s	1.71
Assigned value, μ	54.2
Recovery, %	99.3
$t = \sqrt{p} \cdot (m - \mu) / s$	-1.2014
Sign. level, p(t)	0.2409

No test statistics were found to be significant  
UC denotes a Cochran outlier  
UG denotes a Grubbs outlier

pH,  
Control of recovery, average of results

Laboratory	Sample pair AB	
1	7.490	
2	7.050	
3	7.280	
4	7.010	
5	7.245	
6	7.325	
7	7.245	
8	-	
9	7.270	
10	7.160	
11	7.220	
12	7.675	
13A	7.460	
14	6.900	
15	8.135	UG
16	7.425	
17	7.060	
18	7.300	
19	7.050	
20	-	
21	7.325	
22	7.255	
23	7.185	
24	7.155	
25	7.150	
26A	7.360	
26B	-	
27	6.955	
28	7.000	
29	7.545	
30	7.415	
31	7.175	
32	7.360	
33	7.800	
34	7.050	
35	7.130	
36	7.210	
37	7.032	
38	7.335	
39	6.990	
41	7.375	
42	6.945	
43	7.100	
44	7.130	
45	-	
46	6.245	UC
47	6.985	
48A	7.305	
48B	7.290	
49A	7.055	
49B	-	
50	7.285	
51	7.845	
52A	-	
52B	7.560	
53	7.155	

Laboratory	Sample pair AB	
54	7.395	
55	6.300	UG
56	7.165	
57	6.960	
58	7.215	
59	7.285	
60	7.215	
61	7.216	
62	7.300	
63	7.290	
13B	-	
No of labs., p	57	
No of repl., n	2	
m	7.240	
s <sup>2</sup>	0.040	
s	0.201	
Assigned value, μ	7.24	
Recovery, %	100.0	
$t = \sqrt{p} \cdot (m-\mu)/s$	-0.0112	
Sign. level, p(t)	0.9911	

No test statistics were found to be significant  
UC denotes a Cochran outlier  
UG denotes a Grubbs outlier

## ANNEX E CONCENTRATION LEVEL

### Concentration level SPIL-4 (2015)

Parameter	Unit	Sample	Bottle no.	I	II	Bottle Average	Sample	Assigned	Spike	
							Average	value	Measured	Assigned
Total nitrogen	mg/L N	A1	13	4,08	4,11	4,10	4,08	4,00	-0,73	-0,80
			36	4,15	3,96	4,06				
			63	3,99	4,16	4,08				
		B1	8	4,68	4,81	4,75	4,81	4,8		
			33	5,01	4,79	4,90				
			78	4,77	4,79	4,78				
Ammonium	mg/L N	A1	13	0,53	0,53	0,53	0,53	0,57	-0,13	-0,13
			36	0,53	0,53	0,53				
			63	0,53	0,53	0,53				
		B1	8	0,65	0,66	0,66	0,66	0,70		
			33	0,66	0,66	0,66				
			78	0,66	0,66	0,66				
Nitrate	mg/L N	A1	13	2,08	2,12	2,10	2,11	2,13	-0,55	-0,57
			36	2,13	2,12	2,13				
			63	2,09	2,12	2,11				
		B1	8	2,66	2,66	2,66	2,66	2,7		
			33	2,65	2,65	2,65				
			78	2,69	2,66	2,68				
pH		A2	15	7,26	7,26	7,26	7,25	7,24	0,01	0,00
			23	7,22	7,19	7,21				
			65	7,25	7,30	7,28				
		B2	18	7,24	7,25	7,25	7,24	7,24		
			39	7,22	7,24	7,23				
			79	7,24	7,24	7,24				

## ANNEX F HOMOGENEITY AND STABILITY

SPIL-4
PT:2015
Parameter:NH4
Unit:mg/L N
Sigma:0,0442 6,5% of x

Responsible for tests: SMN/IRL

6,5% level or  $1,3 \cdot S_{T \max}$

Homogeneity test Date: 2015-10-28

Sample	x(a)	x(b)	average	sd	sd^2
B1-5	0,673	0,675	0,7	0,001	0,000
B1-10	0,682	0,683	0,7	0,001	0,000
B1-20	0,678	0,685	0,7	0,005	0,000
B1-23	0,688	0,682	0,7	0,004	0,000
B1-34	0,676	0,682	0,7	0,004	0,000
B1-38	0,681	0,675	0,7	0,004	0,000
B1-49	0,673	0,676	0,7	0,002	0,000
B1-52	0,682	0,681	0,7	0,001	0,000
B1-64	0,687	0,679	0,7	0,006	0,000
B1-70	0,678	0,676	0,7	0,001	0,000
B1-75	0,677	0,681	0,7	0,003	0,000
B1-79	0,678	0,682	0,7	0,003	0,000

### For homogeneity

General average (x)	0,68
Sample average sd ( $s_x$ )	0,003
Within-sample sd ( $s_w$ ):	0,003
Between-samples sd (ss):	0,0024
$S_L$ in the Proficiency Test:	
$S_R$ in the Proficiency Test:	

### Conclusions

	ss =0,002	0.3*sigma=0,01
	x-y  =0	
<b>Analytical quality</b>	Is $s_w < 0,15 \cdot \sigma$ <b>YES</b>	
<b>Homogeneity:</b>	Is ss < 0.3*sigma? <b>YES</b>	

PT:SPIL-4 2015  
 Parameter:NO3  
 Unit:mg/L N  
 Sigma:0,2 1,3\*S<sub>T max</sub>

Responsible for tests: SMN/IRL

6,5% level or 1,3\*S<sub>T max</sub>

Homogeneity test Date:2015-10-28

Sample	x(a)	x(b)	average	sd	sd^2
B1-5	2,65	2,69	2,7	0,028	0,001
B1-10	2,68	2,73	2,7	0,035	0,001
B1-20	2,66	2,7	2,7	0,028	0,001
B1-23	2,70	2,73	2,7	0,021	0,000
B1-34	2,67	2,73	2,7	0,042	0,002
B1-38	2,65	2,67	2,7	0,014	0,000
B1-49	2,70	2,69	2,7	0,007	0,000
B1-52	2,68	2,7	2,7	0,014	0,000
B1-64	2,64	2,65	2,6	0,007	0,000
B1-70	2,67	2,66	2,7	0,007	0,000
B1-75	2,67	2,66	2,7	0,007	0,000
B1-79	2,69	2,64	2,7	0,035	0,001

**For homogeneity**

General average (x) 2,7  
 Sample average sd (s<sub>x</sub>) 0,021  
 Within-sample sd (s<sub>w</sub>): 0,024  
 Between-samples sd (ss): 0,013  
 S<sub>L</sub> in the Proficiency Test:  
 S<sub>R</sub> in the Proficiency Test:

Stability test Date:2015-11-12

Sample	x(a)	x(b)
B1-8	2,66	2,66
B1-33	2,65	2,65
B1-78	2,69	2,66

**For stability**

General average (y): 2,661667  
 /x-y/ = 0,017917

**Conclusions**

ss =0,01	0.3*sigma=0,05
/x-y/ =0,017917	
<b>Analytical quality</b>	Is s <sub>w</sub> < 0,15*sigma <b>YES</b>
<b>Homogeneity:</b>	Is ss < 0.3*sigma? <b>YES</b>
<b>Stability:</b>	/x-y/ < 0.3*sigma? <b>YES</b>



PT:SPIL-4 2015  
 Parameter:pH  
 Unit:  
 Sigma:0,065 6,5% of x

Responsible for tests: SMN/IRL

6,5% level or  $1,3 \cdot S_{T \max}$

Homogeneity test Date:2015-10-28

Sample	x(a)	x(b)	average	sd	sd^2
B2-4	7,42	7,49	7,5	0,049	0,002
B2-9	7,41	7,45	7,4	0,028	0,00
B2-19	7,44	7,46	7,5	0,014	0,000
B2-22	7,42	7,51	7,5	0,064	0,004
B2-34	7,44	7,44	7,4	0,000	0,000
B2-43	7,44	7,43	7,4	0,007	0,000
B2-51	7,47	7,47	7,5	0,000	0,000
B2-53	7,47	7,48	7,5	0,007	0,000
B2-57	7,49	7,45	7,5	0,028	0,001
B2-65	7,46	7,45	7,5	0,007	0,000
B2-75	7,45	7,46	7,5	0,007	0,000
B2-80	7,44	7,44	7,4	0,000	0,000

**For homogeneity**

General average (x) 7,45  
 Sample average sd ( $s_x$ ) 0,015  
 Within-sample sd ( $s_w$ ): 0,027  
 Between-samples sd (ss): 0  
 $S_L$  in the Proficiency Test:  
 $S_R$  in the Proficiency Test:

**Conclusions**

	ss =0	0.3*sigma=0,02
	x-y  =0	
<b>Analytical quality</b>	Is $s_w < 0,15 \cdot \sigma$ <b>NO</b>	
<b>Homogeneity:</b>	Is ss < 0.3*sigma? <b>YES</b>	