# **Certified Reference Values**





Production date: 01-07-2024 Batch: 27 - 2024

		Unhomogenised past. milk	Unhomogenised past. cream										
	Туре	Buttermilk	6,5%	8%	9,5%	11%	12%	14%	18%	32%	38%	46%	50%
Fat (g/100 g) ≤8%+9,5%: ISO 1211/IDF 1:2010 ≥9%: ISO 2450/IDF 16: 2008		0,618	6,429	7,996	9,296	10,771	11,909	15,394**	18,553**	31,894	38,693	44,746	49,290
Protein (g/ 100 g) ISO 8968-1/IDF 20-1:2014		2,814	3,339	3,303	3,243	3,185	3,298	3,216	3,060	2,498	2,251	1,948	1,778
Lactose, Mono (g/100 g) ISO 5765-2/IDF 79:2002											3,118		
Dry Matter (g/100 g) ISO 6731/IDF 21:2010				16,619	17,812		20,169	22,104	25,671	38,400	44,532	49,922	54,267

** Increase uncertainty (U×2)			

Date: 04-07-2024

The results apply only the received batch

CRM Approver, Bente Østergaard

The results may not be reproduced, except in their entirety, without written permission from the Laboratory.

Contact: calibration@ftn.eurofins.com

Information, if any:		

### Informations according to the certified reference material:

#### Description

Fresh Bovine Milk from Danish Dairies. The Samples are added preservative.

The certified values are calulated by accredited reference methods.

Because of the short validy period, the reference value is done by one laboratory, but in several determinations and batch.

Bias are checked by comparing with FTIR values from

Eurofins Milk Testing.

Homogenity are checked on Reference values and FTIR

values.
Stability has been calculated for a representative batch.

## The Reference Laboratory have some approved modification:

Fat: Drying fixed time instead of to constant weight

Protein: The destruction don't start with low temperature, but starts at 420°C Dry Matter: Using disposble cups without lids instead of glass bowls with lids.

The Samples are calculated with following expanded uncertanity U%

	U%
Fat	1,5%
Protein	1,5%
Lactose	2,0%
Dry Matter	1,0%

#### Instruction for handling and use:

The Samples are set at room temperature for 60 min, then heated in water bath to 38°C for 20 min. Gently invert the Samples 20 times and measure immediately on FTIR equipment The Samples can't be reused

## Sample size:

Tube with 50mL

## Period of validity:

14 days at 4°C ± 2°C from production date.

## Storage information:

Store at  $4^{\circ}$ C  $\pm 2^{\circ}$ C until use.